




# Test Certificate

Part 1 Liter Vessel  
Part Number 02503  
Serial Number 91401 - 91505



		USP / Ph.Eur. Tolerances		
Dimensions	Vessel	Inside Ø a	100 mm	98 – 106 mm
		Height b	200 mm	160 – 210 mm
		With hemispherical bottom		
Material	Borosilicate glass			

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at date of approval, see below).

Aesch, 2022-10-13

Bruno Balli  
SOTAX Quality Management  
SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland