



Part 100 mL Cylinder Holder, complete

Part Number 10758



# Test Certificate

|           |                 |      | USP / Ph.Eur.<br>Tolerances |
|-----------|-----------------|------|-----------------------------|
| Parameter | Cylinder Holder | Mass | 240 g ± 12 g                |

The accuracy of the above mentioned part parameters have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The parameters are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at date of approval, see below).

Aesch, 2021-11-09

Bruno Balli  
SOTAX Quality Management  
SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland