

USP/Ph.Eur.

Test Certificate

Part: Paddle with Stirring Shaft

variable for Paddle over Disc

Number: 11365

Serial Number: N170001 – N170040

For SOTAX AT Dissolution Testing Systems



Tolerances Dimensions: Paddle Height 19.0 mm ± 0.5 mm Upper length 74.5 mm ± 0.5 mm Lower length 42.0 mm ± 1.0 mm Size 4.0 mm ± 1.0 mm Radius 1.2 mm ± 0.2 mm Radius 41.5 mm ± 1.0 mm Shaft Diameter 9.4 to 10.1 mm

Material: Stainless Steel (electrolytic polished)

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management Bruno Balli 2017-06-27

Signature

CT11365EN 1 / 1