

## **Test Certificate**

Part:

## Paddle with Stirring Shaft variable for Paddle over Disc

Number:

11365

Serial Number:

N170041 - N170100

For SOTAX AT Dissolution Testing Systems

|             |        | /  |  |  |  |
|-------------|--------|--|--|--|--|
|             |        | L  |  | USP/Ph.Eur.<br>Tolerances  |  |
| Dimensions: | Paddle | Height<br>Upper length<br>Lower length<br>Size<br>Radius<br>Radius | 19.0 mm<br>74.5 mm<br>42.0 mm<br>4.0 mm<br>1.2 mm<br>41.5 mm | ± 0.5 mm<br>± 0.5 mm<br>± 1.0 mm<br>± 1.0 mm<br>± 0.2 mm<br>± 1.0 mm |  |
|             | Shaft  | Diameter   | 9.4 to 10.1 mm   |  |  |
| Material:   |        | Stainless Steel (electrolytic polished)                            |  |  |  |

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management Bruno Balli 2017-07-13

Signature