


Test Certificate

Part: **1 Liter Vessel**
 Number: 14143-02
 Serial Number: 74538 - 74579

For SOTAX AT Dissolution Testing Systems



| | | | | |
|-------------|----------|---|-------------------|---------------------------|
| Dimensions: | Inside Ø | a | 102.9 to 105.1 mm | USP/Ph.Eur. Tolerances |
| | Height | b | 200 mm | |

With hemispherical bottom

Material: Borosilicate glass inactinic brown

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management
 Bruno Balli
 2018-05-24


 Signature